


ragané

CERTIFICATE OF ANALYSIS

Chromatogram

mV



Quantitative Results

Detector A

Compound Name	Concentration, %
CBDV	0.094
CBDA	--
CBGA	--
CBG	1.051
CBD	30.778
THCV	--
CBN	0.486
THC	--
CBC	--
THCA-A	--
CBL	--
CBDVA	--
CBDB	--

Sample information

Sample name: Hemp Drops 3000mg CBD DC
Analysis date: 2022 09 05

Summary

Total THC	0.00	%
Total THC	0.00	mg/g
Total CBD	30.78	%
Total CBD	307.78	mg/g

-- — compound below detection limit or not detected; LOQ <0.01%
THC content does not exceed legal limits.

Instrumental and analytical conditions.

Equipment: Quantitative analysis was performed using Shimadzu Cannabis Analyzer for Potency - an integrated HPLC system with built-in sample cooler, degasser, autoinjector and UV detector. NexLeaf CBX for potency, 2.7 μ m, 4.6 x 150 mm column coupled with NexLeaf CBXGuard column. Shimadzu LabSolutions software.